

### Highlights:

- Results in 5 minutes or less
- Available in 100-strip individual kit format or bulk packaging

### Contents of Kit:

- 100 QuickStix Strips packed in two moisture-resistant canisters
- 100 Disposable Tissue Extractors, each consisting of a tube with punch cap and pestle (optional item with bulk packaging)
- 1 dropper bottle for dispensing tap water



Obtain Leaf Tissue



Grind Tissue

Catalog Number AS 010 LS

## Intended Use

The EnviroLogix QuickStix Kit for Roundup Ready® Plant Tissue screens for the presence of CP4 EPSPS protein in **Corn Leaf, Corn Seed, and Soy Leaf** tissues. It is not currently validated for any other crops. The QuickStix will not detect GA21 Roundup Ready.

To test soybean or corn bulk grain, order Catalog Number AS 010 BG. If interested in testing canola leaf and seed, order Catalog Number AS 017 LS.

## How the Test Works

Crops that have been genetically modified with Roundup Ready® express CP4 EPSPS protein in their tissue. To detect the CP4 EPSPS protein with the EnviroLogix QuickStix Strip, tissue samples must be extracted and the protein solubilized in tap water.

Each QuickStix Strip has an absorbent pad at each end. The protective tape with the arrow indicates which end of the strip to insert into the extraction tube. The sample travels up the membrane strip and is absorbed into the larger pad at the top of the strip. The portion of the strip between the protective tape and the absorbent pad at the top of the strip is used to view the reactions as described under “Interpreting the Results”.

## Sample Preparation

Prior to beginning, fill the dropper bottle provided with water.

### To extract corn leaf tissue:

1. Sandwich a section of leaf tissue between the cap and body of the Disposable Tissue Extractor tube. Snap **one** circular tissue punch by closing the cap. Push the leaf punch down into the tapered bottom of the tube with the pestle. Write the sample identification on the tube with a waterproof marker.
2. Insert the pestle into the tube and grind the tissue by rotating the pestle against the sides of the tube with twisting motions. Continue this process for 20 to 30 seconds, or until the leaf tissue is well ground.
3. Holding the dropper bottle vertically, carefully add 10 drops of water into each tube containing corn leaf tissue. If using a pipette, add 500 µL of water.
4. Repeat the grinding step to mix tissue with water. Dispose of the pestle. Do not re-use pestles on more than one sample.

### To extract corn seed:

1. Crush a single corn seed (*Suggestion: use pliers with seed in microcentrifuge tube or resealable plastic bag*). Transfer to a tissue extraction tube marked with sample identification. Note: Complete crushing of seed improves extraction efficiency and test performance.
2. Holding the dropper bottle vertically, add 20 drops of water into the tube. If using a pipette, add 1 mL of water.
3. Close the tube cap securely. Shake the tube vigorously for 20 to 30 seconds. Allow the solid material to settle to the bottom of the tube.



Crush Seeds



Extract Seed Sample



Insert QuickStix Strip



Any clearly discernable pink Test Line is considered positive

- Repeat the protocol for each sample to be tested, using a new tube for each. Use caution to prevent sample-to-sample cross-contamination with plant tissue, fluids, crushing equipment (*pliers*) or disposables.

### To extract soy leaf tissue:

*Note: Tissue from leaf sets 1-4 or cotyledons should be selected when testing plants younger than 36 days. Since soy leaves are high expressers of CP4 EPSPS protein take care to avoid cross-contamination when sampling the leaf tissue. Plant tissue older than 36 days should be sampled from cotyledon or leaf set 1 or 2 only.*

- Sandwich a section of tissue between the cap and body of the Disposable Tissue Extractor tube.
  - For leaf tissue, snap **one** circular tissue punch by closing the cap. The soy leaves lowest on each plant (first or second set above the cotyledons), should be used.
  - For cotyledons, snap  $\frac{1}{2}$  circular tissue punch by placing tissue halfway across the tube before closing the cap.
- Push the tissue punch down into the tapered bottom of the tube with the pestle. Write the sample identification on the tube with a waterproof marker. **Do not grind the tissue.** Do not re-use pestles on more than one sample.
- Holding the dropper bottle vertically, carefully add 10 drops of water into each tube containing unground leaf or cotyledon tissue. If using a pipette, add 500  $\mu$ L of water.
- Close the tube cap securely. Shake the tube vigorously for 20 – 30 seconds. Allow the solid material to settle to the bottom of the tube.

## How to Run the QuickStix Strip Test


- Allow refrigerated canisters to come to room temperature before opening. Remove the QuickStix Strips to be used. Avoid bending the strips. Reseal the canister immediately.
- Place the strip into the extraction tube. The sample will travel up the strip. Use a rack to support multiple tubes if needed.
- Allow the strip to develop for 5 minutes before making final assay interpretations. Positive sample results may become obvious much more quickly.
- To retain the strip, cut off and discard the bottom section of the strip covered by the arrow tape.

## Interpreting the Results

Development of the Control Line within 5 minutes indicates that the strip has functioned properly. Any strip that does not develop a Control Line should be discarded and the sample re-tested using another strip.

If the sample extract contains the CP4 EPSPS protein, a second line (Test Line) will develop on the membrane strip between the Control Line and the protective arrow tape. The results should be interpreted as positive for CP4 EPSPS protein expression.

If no Test Line is observed after 5 minutes, the results should be interpreted as negative. A negative result means the sample contains less CP4 EPSPS than is typically expressed in the tissues of Roundup Ready plants.





## Kit Storage

This QuickStix Kit can be stored at room temperature, or refrigerated for a longer shelf life. Note the shelf life on the kit box for each storage temperature. The kit may be used in field applications; however, prolonged exposure to high temperatures may adversely affect the test results. Do not open the desiccated canister until ready to use the test strips.

## Precautions and Notes

- This kit is designed for screening for presence or absence only and is not meant to be quantitative.
- As with all tests, it is recommended that results be confirmed by an alternate method when necessary.
- The assay has been optimized to be used with the protocol provided in the kit. Deviation from this protocol may invalidate the results of the test.
- The results generated through the proper use of this kit reflect the condition of the working sample directly tested. Extrapolation as to the condition of the originating lot from which the working sample was derived should be based on sound sampling procedures and statistical calculations which address random sampling effects, non-random seed lot sampling effects, and assay system uncertainty. A negative result obtained when properly testing the working sample does not necessarily mean the originating lot is entirely negative for the analyte or protein in question.
- A negative result with this kit does not mean that the sampled tissue has not been otherwise genetically modified.
- A strong positive result may safely be interpreted in as little as 2 minutes after sample addition. It is not safe, however, to conclude that a sample is negative before a full 5 minutes has elapsed. A weakly positive sample may require the full 5 minutes for a distinct Test Line to appear.
- Protect all components from hot or cold extremes of temperature when not in use. Do not leave in direct sunlight or in vehicle.



**For Technical Support  
Contact Us At:**

**EnviroLogix**

500 Riverside Industrial  
Parkway  
Portland, ME 04103-1486  
USA

**Tel: (207) 797-0300**

**Toll Free: 866-408-4597**

**Fax: (207) 797-7533**

e-mail:

***info@envirologix.com***

website:

***www.envirologix.com***



## LIMITED WARRANTY

EnviroLogix Inc. ("EnviroLogix") warrants the products sold hereunder ("the Products") against defects in materials and workmanship when used in accordance with the applicable instructions for a period not to extend beyond a product's printed expiration date. If the Products do not conform to this Limited Warranty and the customer notifies EnviroLogix in writing of such defects during the warranty period, including an offer by the customer to return the Products to EnviroLogix for evaluation, EnviroLogix will repair or replace, at its option, any product or part thereof that proves defective in materials or workmanship within the warranty period.

**NEITHER ENVIROLOGIX NOR MONSANTO MAKES ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.** The warranty provided herein and the data, specifications and descriptions of EnviroLogix products appearing in EnviroLogix published catalogues and product literature are EnviroLogix' sole representations concerning the Products and warranty. No other statements or representations, written or oral, by EnviroLogix' employees, agents or representatives, except written statements signed by a duly authorized officer of EnviroLogix Inc., are authorized; they should not be relied upon by the customer and are not a part of the contract of sale or of this warranty.

EnviroLogix does not warrant against damages or defects arising in shipping or handling, or out of accident or improper or abnormal use of the Products; against defects in products or components not manufactured by EnviroLogix, or against damages resulting from such non-EnviroLogix made products or components. EnviroLogix passes on to customer the warranty it received (if any) from the maker thereof of such non-EnviroLogix made products or components. This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by EnviroLogix.

**THIS WARRANTY IS EXCLUSIVE.** The sole and exclusive obligation of EnviroLogix shall be to repair or replace the defective Products in the manner and for the period provided above. EnviroLogix shall not have any other obligation with respect to the Products or any part thereof, whether based on contract, tort, strict liability or otherwise. Under no circumstances, whether based on this Limited Warranty or otherwise, shall EnviroLogix be liable for incidental, special, or consequential damages.

This Limited Warranty states the entire obligation of EnviroLogix with respect to the Products. If any part of this Limited Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

## License

EnviroLogix has developed this kit using proprietary reagents as well as reagents licensed from Monsanto Company.

*Roundup Ready is a registered trademark of Monsanto Technology, LLC  
EnviroLogix, the EnviroLogix logo, QuickStix and Common Extraction are trademarks of EnviroLogix Inc.*

© EnviroLogix 2007